



World Precision Instruments, Inc.
175 Sarasota Center Boulevard
Sarasota, FL 34240



140 Hale Street
Haverhill, MA 01830
qc@embryotech.com

ELI Accession Number: WPI-2282-0122

Date of completion: 07-17-2023

Lot number: GCD50793

Reference number: W041519

Description of test article: KWIK-SIL

Assay system requested by customer: The test article was ejected through a static mixing tip onto a 35 mm dish to form the walls of a well. The adhesive was allowed to dry for 30 minutes, disinfected with 70% Isopropyl Alcohol, then rinsed with culture medium. 700 μ l of culture medium was then added to the well. One-cell embryos were then added to the well and culture for 96-hours.

Control assay method and results: 21 one-cell (B6D2F1 X B6C3F1) embryos were cultured in triplicate micro drops of culture medium and overlaid with oil:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 80% of one-cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 one-cell (B6D2F1 X B6C3F1) embryos were cultured in the test article using culture medium:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr


Pass/Fail = Pass

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos and were cultured in the same incubator at 37°C and 5.0% CO₂. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the well formed by the test article developed to the expanded blastocyst stage within 96-hours.


Signature
Study Director


Date


Signature
Quality Reviewer


Date